

University of Groningen

**In situ monitoring of polymer redox states by resonance  $\mu$ Raman spectroscopy and its applications in polymer modified microfluidic channels**

Logtenberg, Hella; Jellema, Laurens-Jan C.; Lopez-Martinez, Maria J.; Areephong, Jetsuda; Verpoorte, Elisabeth; Feringa, Ben L.; Browne, Wesley R.

*Published in:*  
Analytical Methods

*DOI:*  
[10.1039/C1AY05475A](https://doi.org/10.1039/C1AY05475A)

**IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.**

*Document Version*  
Publisher's PDF, also known as Version of record

*Publication date:*  
2012

[Link to publication in University of Groningen/UMCG research database](#)

*Citation for published version (APA):*

Logtenberg, H., Jellema, L.-J. C., Lopez-Martinez, M. J., Areephong, J., Verpoorte, E., Feringa, B. L., & Browne, W. R. (2012). In situ monitoring of polymer redox states by resonance  $\mu$ Raman spectroscopy and its applications in polymer modified microfluidic channels. *Analytical Methods*, 4(1), 73-79.  
<https://doi.org/10.1039/C1AY05475A>

**Copyright**

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

**Take-down policy**

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.


## Electronic Supplementary Information

# In situ monitoring of polymer redox states by Resonance microRaman spectroscopy and its applications in polymer modified microfluidic channels

5 Hella Logtenberg,<sup>a</sup> Laurens-Jan Jellema,<sup>b</sup> Maria J. Lopez-Martinez,<sup>b</sup> Jetsuda Areephong,<sup>a</sup> Elisabeth Verpoorte,<sup>b</sup> Ben L. Feringa<sup>a</sup> and Wesley R. Browne<sup>\*a</sup>

10


Electropolymerisation on gold bead electrodes was performed to study the scan rate dependence with respect to film thickness. Different film thicknesses lead to different electrochemical scan rate dependent responses. Charge trapping is commonly observed for thicker polymer films, as described schematically in Fig S 1.




15

**Fig S 1** Trapped charges depend on polymer film thickness.

The cyclic voltammograms from which the scan rate dependence was obtained are depicted in Fig S 2.



**Fig S 2** Cyclic Voltammetry of a thick polymer film on a gold bead electrode (scan rate is in  $\text{V s}^{-1}$ )




20 **Fig S 3** Optical microscope pictures of a thin (left) and thick (right) polymer film on a gold bead electrode, 100x magnification. On the right aggregates of polymer are observed as darker spots.

A model system for this sexithiophene polymer has been reported previously, which is a phenyl capped dimer.<sup>1</sup> For further evidence of the monocationic species observed in Raman spectroscopy this dimer was also studied.


**Fig S 4** Structure of model compound, phenyl capped dimer.

5 The monocation can be formed in solution by addition of TFA or Fe(III)Cl<sub>3</sub>. The formation of the monocation is confirmed by UV/Vis spectroscopy. Resonance enhanced Raman signals are obtained corresponding to the spectrum of monopolaron of the polymer (Fig S 5).




**Fig S 5** Raman signals obtained by adding Fe(III)Cl<sub>3</sub> to a solution of the model compound 2.

10




**Fig S 6** Raman spectrum of polymer on ITO electrode at  $\lambda_{\text{exc}}$  633 nm.




**Fig S 7** Raman spectrum of thin polymer film on a gold bead electrode at  $\lambda_{\text{exc}}$  532 nm

15




**Fig S 8** Raman spectroelectrochemistry on a thin polymer film on a gold bead electrode recorded at  $\lambda_{\text{exc}}$  532 nm.

An overview of the set-up used for electrochemistry in the microchannel is depicted in Fig S 9.



**Fig S 9** Microfluidic channel used for polymerisation.



**Fig S 10** Polymer was only observed on the gold-electrode connected to the potentiostat as working electrode, which was confirmed by Raman spectroscopy ( $\lambda_{\text{exc}}$  785 nm), upper = left electrode, modified with polymer, lower = right electrode, not modified, Raman signals of PDMS observed.

<sup>1</sup> J. Areephong, J. H. Hurenkamp, M. T. W. Milder, A. Meetsma, J. L. Herek, W. R. Browne and B. L. Feringa, *Org. Lett.*, 2009, **11**, 721; M. T. W. Milder, J. L. Herek, J. Areephong, B. L. Feringa and W. R. Browne, *Chem. Phys. Lett.*, 2009, **479**, 137